

**ANNEXURE A  
SCHEDULE OF ACCREDITATION**

**MASS METROLOGY**

**Laboratory Accreditation Number: CAL-8 002 (ISO/IEC 17025:2017)**

| <p><b>Permanent Address of Laboratory</b><br/>Namibian Standards Institution<br/>Metrology Department<br/>205 Gold Street, Prosperita<br/>Windhoek<br/>Namibia</p> <p><b>Postal Address</b><br/>P O Box 26364<br/>Windhoek<br/>Namibia</p> <p><b>Tel</b> : +264 61 386 470/481<br/><b>Cell</b> : +264 81 261 3694<br/><b>Fax</b> : +264 61 386 477<br/><b>Email</b> : <a href="mailto:matalis@nsi.com.na">matalis@nsi.com.na</a></p> |   | <p><b>Technical Signatories</b> : Mr S Matali (All Items)<br/>Mr SS Sankwasa (All Items)<br/>Ms Lusia Katanga (All Items)<br/>Ms P Sheehama (Mass pieces<br/>1 mg to 5 kg)</p> <p><b>Nominated Representative</b> : Mr S Matali</p> <p><b>Issue No</b> : 05<br/><b>Date of Issue</b> : 14 January 2022<br/><b>Expiry Date</b> : 28 February 2023</p> |  |  |
|--|---|--|--|--|
| ITEM   | MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT    | METHOD   | RANGE OF MEASURED QUANTITY   | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)                                 |
|  |   |  | <b>At NSI</b>  |  |
| 1  | Mass pieces   | Internal: MTPI 002<br><br>Reference: OIML R111-1   | 1 mg to 5 g<br>10 g<br>20 g to 50 g<br>100 g<br>200 g<br>500 g<br>1 kg to 2 kg<br>5 kg<br>10 kg<br>20 kg | 0,02 mg<br>0,03 mg<br>0,04 mg<br>0,1 mg<br>0,2 mg<br>0,001 g<br>0,002 g<br>0,005 g<br>0,03 g<br>0,03 g |
|  |   |  | <b>At NSI and On-site</b>  |  |
| 2  | Weighing Instruments<br>Digital – self - indicating | Internal: MTPI 003<br><br>Reference: OIML R76, EURAMET cg 18   | 0 g to 50 g<br>50 g to 200 g<br>200 g to 3000 g<br>3,0 kg to 1500 kg                                     | 0,00002 g<br>0,0006 g<br>0,00015 % + 0,00 g<br>0,001%  |

Original date of accreditation: 20 February 2013

Page 1 of 1

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor  $k = 2$ , corresponding to a confidence level of approximately 95%.



**Pinkie J Malebe**  
**SADCAS Technical Manager**