

**ANNEXURE A
SCHEDULE OF ACCREDITATION**

MASS METROLOGY

Laboratory Accreditation Number: CAL-8 002 (ISO/IEC 17025:2017)

<p>Permanent Address of Laboratory Namibian Standards Institution Metrology Department 205 Gold Street, Prosperita Windhoek Namibia</p> <p>Postal Address P O Box 26364 Windhoek Namibia</p> <p>Tel : +264 61 386 470/481 Cell : +264 81 261 3694 Fax : +264 61 386 477 Email : matalis@nsi.com.na</p>		<p>Technical Signatories : Mr S Matali (All Items) Mr SS Sankwasa (All Items) Ms L Katanga (All Items) Ms P Sheehama (Mass pieces 1 mg to 5 kg)</p> <p>Nominated Representative : Mr S Matali</p> <p>Issue No : 01 Date of Issue : 28 February 2023 Expiry Date : 27 February 2028</p>		
ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	METHOD	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)
			At NSI	
1	Mass pieces	Internal: MTPI 002 Reference: OIML R111-1	1 mg to 5 g 10 g 20 g to 50 g 100 g 200 g 500 g 1 kg to 2 kg 5 kg 10 kg 20 kg	0,02 mg 0,03 mg 0,04 mg 0,1 mg 0,2 mg 0,001 g 0,002 g 0,005 g 0,03 g 0,03 g
			At NSI and On-site	
2	Weighing Instruments Digital – self - indicating	Internal: MTPI 003 Reference: OIML R76, EURAMET cg 18	0 g to 50 g 50 g to 200 g 200 g to 3000 g 3,0 kg to 1500 kg	0,00002 g 0,0006 g 0,00015 % + 0,00 g 0,001%

Original date of accreditation: 20 February 2013

Page 1 of 1

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%.


Pinkie J Malebe
SADCAS Technical Manager