

ANNEXURE A

SCHEDULE OF ACCREDITATION

VOLUME METROLOGY


Laboratory Accreditation Number: **CAL-9 003 (ISO/IEC 17025:2017)**

<p>Permanent Address of Laboratory Namibian Standards Institution Metrology Department 205 Gold Street, Prosperita Windhoek Namibia</p> <p>Postal Address P O Box 26364 Windhoek Namibia</p> <p>Tel : +264 61 386 470/481 Cell : +264 81 261 3694 Fax : +264 61 386 477 Email : matalis@nsi.com.na</p>		<p>Technical Signatories : Mr S Matali (All Items) Mr S S Sankwasa (All Items) Ms P Sheehama (Micro-pipettes only)</p> <p>Nominated Representative : Mr S Matali</p> <p>Issue No : 01 Date of Issue : 28 February 2023 Expiry Date : 27 February 2028</p>		
ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	METHOD	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)
			At NSI	
1	Micropipettes / Syringes	Internal: MTPI 009 Reference: ISO 8655-6	1 μl to 10 μl 10 μl to 100 μl 100 μl to 200 μl 200 μl to 500 μl 500 μl to 1000 μl	0,2 μl 0,8 μl 0,9 μl 1,6 μl 8,2 μl
			At NSI	
2	Glassware	Internal: MTPI 010 & MTPI 011 Reference: ISO 4787	10 ml to 5 l	0,02 %
			At NSI	
3	Metal Measures	Internal: MTPI 010 & MTPI 014 Reference: ISO 4787 OIML R120	1000 ml to 20 l	0,04 %

Original date of accreditation: 20 February 2013

Page 1 of 1

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%.



Pinkie J Malebe
SADCAS Technical Manager